


VERIFICATION OF CALIBRATION REPORT

OCT 26 2010

of DataMaster cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130136 ✓
Name		STEVEN C. KEVAN	ID# 6585 ✓ Date 10/14/10 ✓
A	Agency	AST	Phone # 543-2294
Instrument Location		BETHEL AST	
Alco S/N		X 124623 ✓	Target Value .081 ✓ High Pressure 1250
B	Alco Test Values		.084 ✓
	1 st Alco		2 nd Alco
Signature			
(OVER)			


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.



Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of DEC, 2010.


Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)

BT9 06/03



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130136

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130136

OCTOBER 14, 2015

OPERATOR'S NAME:
KEVIN / STEVE
OPERATOR'S NUMBER: 6585
SUBJECT'S LAST NAME:
TEST

SUBJECT'S FIRST NAME/NI :

O.L. #: 00
DEPT/AGENCY: BETH
CASE/REPORT: 00
TEST TYPE: U
ALCO TARGET VALUE: .001
ALCO S/N: X124623

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.81 in
ALCO TARGET .000 17:32
BLANK TEST .000 17:33
INTERNAL STANDARD VERIFIED 17:33
ALCO TO 29.81 in .004 17:34
BLANK TEST .000 17:34
SUBJECT SAMPLE .000 17:35
BLANK TEST .000 17:35
ALCO TO 29.83 in .004 17:36
BLANK TEST .000 17:36

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130136

OCTOBER 14, 2015
TIME 17:30

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 43c
BAROMETER: 29.81 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""#%&'()*+,-./0123456789:;<=>?@AB[CD]EFG
HIJKL NOPQRSTU VWXYZ abcdefghijklmnop
qrstuvwxy z/1234567890